

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036207.D
 Acq On : 19 Jun 2023 15:40
 Operator : JC/MD
 Sample : 03238-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PMW-1

Quant Time: Jun 20 03:43:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	178201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	324534	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159727	49.048	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.100%
35) Dibromofluoromethane	5.385	113	104138	47.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	8.647	98	367871	45.698	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.400%#
62) 4-Bromofluorobenzene	11.079	95	143431	44.140	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.280%
Target Compounds						
16) Acetone	2.380	43	4031	2.979	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036207.D
Acq On : 19 Jun 2023 15:40
Operator : JC/MD
Sample : 03238-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PMW-1

Quant Time: Jun 20 03:43:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

