

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036208.D
 Acq On : 19 Jun 2023 16:04
 Operator : JC/MD
 Sample : 03238-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Jun 20 03:44:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	171665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	311945	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	282849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166920	53.209	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.420%
35) Dibromofluoromethane	5.385	113	109605	51.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	8.647	98	382927	49.488	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.079	95	145569	46.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						Qvalue
16) Acetone	2.386	43	3280	2.517	ug/l	# 84
30) Chloroform	5.099	83	3791	0.873	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036208.D
Acq On : 19 Jun 2023 16:04
Operator : JC/MD
Sample : 03238-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Jun 20 03:44:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

