

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036210.D
 Acq On : 19 Jun 2023 16:51
 Operator : JC/MD
 Sample : 03238-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Jun 20 03:44:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	299557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137099	45.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.980%
35) Dibromofluoromethane	5.385	113	87902	43.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.540%
50) Toluene-d8	8.646	98	305792	41.154	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	82.300%#
62) 4-Bromofluorobenzene	11.079	95	118726	39.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.160%#
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	3076	2.484	ug/l #	88
30) Chloroform	5.098	83	5886	1.426	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036210.D
Acq On : 19 Jun 2023 16:51
Operator : JC/MD
Sample : 03238-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Jun 20 03:44:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

