

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036215.D
 Acq On : 19 Jun 2023 18:49
 Operator : JC/MD
 Sample : PB153531ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#03

Quant Time: Jun 20 03:46:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	274088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	157920	64.502	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 129.000%#		
35) Dibromofluoromethane	5.385	113	99186	53.365	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.720%		
50) Toluene-d8	8.647	98	331159	48.709	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.420%		
62) 4-Bromofluorobenzene	11.079	95	130275	47.470	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.940%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	130319	128.118	ug/l	98
18) Methyl Acetate	2.709	43	5024	1.302	ug/l	94
20) Methylene Chloride	2.788	84	6827	3.704	ug/l	93
25) 2-Butanone	4.568	43	41112	28.870	ug/l	94
37) Ethyl Acetate	4.715	43	52935	16.595	ug/l	99
43) Isopropyl Acetate	6.342	43	209775	42.972	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036215.D
Acq On : 19 Jun 2023 18:49
Operator : JC/MD
Sample : PB153531ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#03

Quant Time: Jun 20 03:46:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

