

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036218.D
 Acq On : 19 Jun 2023 19:59
 Operator : JC/MD
 Sample : PB153531ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#06

Quant Time: Jun 20 03:48:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	132622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	263347	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152425	62.892	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 125.780%#		
35) Dibromofluoromethane	5.385	113	96378	53.969	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 107.940%		
50) Toluene-d8	8.647	98	335426	51.349	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.700%		
62) 4-Bromofluorobenzene	11.079	95	129739	49.203	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.400%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	146473	145.466	ug/l	98
18) Methyl Acetate	2.703	43	4844	1.268	ug/l	97
20) Methylene Chloride	2.788	84	7300	4.001	ug/l	97
25) 2-Butanone	4.562	43	46433	32.939	ug/l	97
37) Ethyl Acetate	4.715	43	52326	17.073	ug/l	99
43) Isopropyl Acetate	6.336	43	206651	44.059	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036218.D
Acq On : 19 Jun 2023 19:59
Operator : JC/MD
Sample : PB153531ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#06

Quant Time: Jun 20 03:48:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

