

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036219.D
 Acq On : 19 Jun 2023 20:23
 Operator : JC/MD
 Sample : PB153531ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#07

Quant Time: Jun 20 03:48:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	266688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244980	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159937	64.764	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 129.520%#		
35) Dibromofluoromethane	5.385	113	99071	54.782	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.560%		
50) Toluene-d8	8.647	98	333265	50.379	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.760%		
62) 4-Bromofluorobenzene	11.079	95	131062	49.082	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.160%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	167390	163.146	ug/l	99
18) Methyl Acetate	2.709	43	5130	1.318	ug/l	98
20) Methylene Chloride	2.788	84	8299	4.464	ug/l	93
25) 2-Butanone	4.562	43	54887	38.211	ug/l	100
37) Ethyl Acetate	4.715	43	52479	16.909	ug/l	99
43) Isopropyl Acetate	6.336	43	211437	44.514	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036219.D
Acq On : 19 Jun 2023 20:23
Operator : JC/MD
Sample : PB153531ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#07

Quant Time: Jun 20 03:48:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

