

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036223.D
 Acq On : 19 Jun 2023 21:56
 Operator : JC/MD
 Sample : PB153531ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#11

Quant Time: Jun 20 03:50:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	269352	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	99314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155269	61.191	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 122.380%#		
35) Dibromofluoromethane	5.385	113	97351	53.299	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.600%		
50) Toluene-d8	8.647	98	322385	48.252	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.500%		
62) 4-Bromofluorobenzene	11.079	95	119915	44.463	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 88.920%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	165626	157.107	ug/l	100
18) Methyl Acetate	2.709	43	5805	1.452	ug/l	97
20) Methylene Chloride	2.794	84	8646	4.527	ug/l	89
25) 2-Butanone	4.562	43	56049	37.976	ug/l	97
37) Ethyl Acetate	4.715	43	54059	17.245	ug/l	100
43) Isopropyl Acetate	6.342	43	212048	44.201	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036223.D
Acq On : 19 Jun 2023 21:56
Operator : JC/MD
Sample : PB153531ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#11

Quant Time: Jun 20 03:50:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

