

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036224.D
 Acq On : 19 Jun 2023 22:20
 Operator : JC/MD
 Sample : PB153531ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#12

Quant Time: Jun 20 03:51:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249130	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127567	52.289	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.580%		
35) Dibromofluoromethane	5.385	113	79570	47.100	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.200%		
50) Toluene-d8	8.647	98	267879	43.349	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 86.700%#		
62) 4-Bromofluorobenzene	11.079	95	105973	42.483	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 84.960%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	163040	160.853	ug/l	98
18) Methyl Acetate	2.703	43	5067	1.318	ug/l	96
20) Methylene Chloride	2.794	84	8622	4.695	ug/l	94
25) 2-Butanone	4.562	43	52457	36.967	ug/l	97
37) Ethyl Acetate	4.708	43	53134	18.326	ug/l	99
43) Isopropyl Acetate	6.342	43	202710	45.685	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036224.D
Acq On : 19 Jun 2023 22:20
Operator : JC/MD
Sample : PB153531ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#12

Quant Time: Jun 20 03:51:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

