

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036225.D
 Acq On : 19 Jun 2023 22:43
 Operator : JC/MD
 Sample : PB153531ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153531ZHE#13

Quant Time: Jun 20 03:51:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146451	73.070	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 146.140%#			
35) Dibromofluoromethane	5.385	113	92295	62.365	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 124.740%#			
50) Toluene-d8	8.647	98	296299	54.735	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 109.460%			
62) 4-Bromofluorobenzene	11.079	95	112374	51.426	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.860%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	159825	191.935	ug/l	100
18) Methyl Acetate	2.709	43	4921	1.558	ug/l	98
20) Methylene Chloride	2.788	84	7946	5.267	ug/l	97
37) Ethyl Acetate	4.715	43	51579	20.308	ug/l	99
43) Isopropyl Acetate	6.336	43	203497	52.354	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036225.D
Acq On : 19 Jun 2023 22:43
Operator : JC/MD
Sample : PB153531ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153531ZHE#13

Quant Time: Jun 20 03:51:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

