


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036198.D
Acq On : 19 Jun 2023 11:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Title : SW846 8260
Last Update : Fri Jun 16 16:28:40 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	14168	PASS
75	95	30	60	57.3	37789	PASS
95	95	100	100	100.0	65949	PASS
96	95	5	9	6.7	4432	PASS
173	174	0.00	2	1.0	394	PASS
174	95	50	100	62.7	41355	PASS
175	174	5	9	7.6	3152	PASS
176	174	95	101	98.2	40595	PASS
177	176	5	9	6.6	2686	PASS