

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
 Data File : VX041937.D
 Acq On : 19 Jun 2024 10:44
 Operator : JC/MD
 Sample : VX0619MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619MBL01

Quant Time: Jun 20 00:55:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149085	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	230756	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75451	43.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.780%
35) Dibromofluoromethane	5.385	113	71846	43.880	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.760%
50) Toluene-d8	8.647	98	270927	47.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.200%
62) 4-Bromofluorobenzene	11.079	95	103345	48.740	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
Data File : VX041937.D
Acq On : 19 Jun 2024 10:44
Operator : JC/MD
Sample : VX0619MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619MBL01

Quant Time: Jun 20 00:55:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

