

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
 Data File : VX041942.D
 Acq On : 19 Jun 2024 12:44
 Operator : JC/MD
 Sample : P2928-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-TB-20240613

Quant Time: Jun 20 00:58:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	165399	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	251454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84604	44.359	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.720%
35) Dibromofluoromethane	5.385	113	79572	44.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.200%
50) Toluene-d8	8.647	98	297511	47.465	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.940%
62) 4-Bromofluorobenzene	11.079	95	109836	47.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
Data File : VX041942.D
Acq On : 19 Jun 2024 12:44
Operator : JC/MD
Sample : P2928-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-TB-20240613

Quant Time: Jun 20 00:58:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

