

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
 Data File : VX041936.D
 Acq On : 19 Jun 2024 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 00:54:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.381	0.336	11.8	85	0.00
3 P	Chloromethane	0.478	0.418	12.6	86	0.00
4 C	Vinyl Chloride	0.479	0.414	13.6#	84	0.00
5 T	Bromomethane	0.237	0.231	2.5	94	0.00
6 T	Chloroethane	0.241	0.198	17.8	91	0.00
7 T	Trichlorofluoromethane	0.628	0.550	12.4	88	0.00
8 T	Diethyl Ether	0.295	0.299	-1.4	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.404	6.5	90	0.00
10 T	Methyl Iodide	0.537	0.495	7.8	89	0.00
11 T	Tert butyl alcohol	0.109	0.111	-1.8	95	0.00
12 CM	1,1-Dichloroethene	0.431	0.389	9.7#	86	0.00
13 T	Acrolein	0.123	0.142	-15.4	110	0.00
14 T	Allyl chloride	0.684	0.684	0.0	94	0.00
15 T	Acrylonitrile	0.286	0.306	-7.0	99	0.00
16 T	Acetone	0.253	0.243	4.0	100	0.00
17 T	Carbon Disulfide	0.879	0.738	16.0	85	0.00
18 T	Methyl Acetate	0.691	0.761	-10.1	98	0.00
19 T	Methyl tert-butyl Ether	1.425	1.514	-6.2	99	0.00
20 T	Methylene Chloride	0.554	0.505	8.8	95	0.00
21 T	trans-1,2-Dichloroethene	0.438	0.407	7.1	90	0.00
22 T	Diisopropyl ether	1.446	1.504	-4.0	97	0.00
23 T	Vinyl Acetate	1.235	1.340	-8.5	98	0.00
24 P	1,1-Dichloroethane	0.810	0.805	0.6	94	0.00
25 T	2-Butanone	0.394	0.418	-6.1	98	0.00
26 T	2,2-Dichloropropane	0.586	0.575	1.9	93	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.548	0.0	94	0.00
28 T	Bromochloromethane	0.347	0.353	-1.7	101	0.00
29 T	Tetrahydrofuran	0.254	0.268	-5.5	96	0.00
30 C	Chloroform	0.843	0.853	-1.2#	96	0.00
31 T	Cyclohexane	0.674	0.596	11.6	87	0.00
32 T	1,1,1-Trichloroethane	0.705	0.678	3.8	90	0.00
33 S	1,2-Dichloroethane-d4	0.577	0.537	6.9	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.355	0.331	6.8	92	0.00
36 T	1,1-Dichloropropene	0.377	0.344	8.8	89	0.00
37 T	Ethyl Acetate	0.491	0.526	-7.1	97	0.00
38 T	Carbon Tetrachloride	0.416	0.392	5.8	89	0.00
39 T	Methylcyclohexane	0.476	0.436	8.4	89	0.00
40 TM	Benzene	1.271	1.226	3.5	93	0.00
41 T	Methacrylonitrile	0.266	0.296	-11.3	101	0.00
42 TM	1,2-Dichloroethane	0.409	0.422	-3.2	97	0.00
43 T	Isopropyl Acetate	0.753	0.813	-8.0	100	0.00
44 TM	Trichloroethene	0.341	0.326	4.4	91	0.00
45 C	1,2-Dichloropropane	0.312	0.322	-3.2#	96	0.00
46 T	Dibromomethane	0.228	0.237	-3.9	97	0.00
47 T	Bromodichloromethane	0.418	0.440	-5.3	97	0.00
48 T	Methyl methacrylate	0.360	0.398	-10.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
 Data File : VX041936.D
 Acq On : 19 Jun 2024 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 00:54:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	97	0.00
50 S	Toluene-d8	1.246	1.137	8.7	89	0.00
51 T	4-Methyl-2-Pentanone	0.519	0.568	-9.4	98	0.00
52 CM	Toluene	0.809	0.800	1.1#	92	0.00
53 T	t-1,3-Dichloropropene	0.409	0.450	-10.0	98	0.00
54 T	cis-1,3-Dichloropropene	0.458	0.499	-9.0	97	0.00
55 T	1,1,2-Trichloroethane	0.337	0.358	-6.2	97	0.00
56 T	Ethyl methacrylate	0.524	0.584	-11.5	98	0.00
57 T	1,3-Dichloropropane	0.545	0.572	-5.0	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.244	0.271	-11.1	97	0.00
59 T	2-Hexanone	0.397	0.440	-10.8	98	0.00
60 T	Dibromochloromethane	0.366	0.403	-10.1	98	0.00
61 T	1,2-Dibromoethane	0.350	0.377	-7.7	100	0.00
62 S	4-Bromofluorobenzene	0.459	0.456	0.7	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.364	0.328	9.9	92	0.00
65 PM	Chlorobenzene	1.040	1.011	2.8	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.373	-2.5	96	0.00
67 C	Ethyl Benzene	1.630	1.586	2.7#	93	0.00
68 T	m/p-Xylenes	0.658	0.640	2.7	93	0.00
69 T	o-Xylene	0.651	0.649	0.3	94	0.00
70 T	Styrene	1.092	1.145	-4.9	96	0.00
71 P	Bromoform	0.326	0.355	-8.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.040	2.857	6.0	92	0.00
74 T	N-amyl acetate	1.350	1.391	-3.0	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.130	1.114	1.4	99	0.00
76 T	1,2,3-Trichloropropane	0.908	0.895	1.4	101	0.00
77 T	Bromobenzene	0.905	0.879	2.9	96	0.00
78 T	n-propylbenzene	3.299	3.203	2.9	94	0.00
79 T	2-Chlorotoluene	2.125	2.017	5.1	95	0.00
80 T	1,3,5-Trimethylbenzene	2.532	2.457	3.0	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.336	-5.0	100	0.00
82 T	4-Chlorotoluene	2.349	2.306	1.8	95	0.00
83 T	tert-Butylbenzene	2.779	2.628	5.4	92	0.00
84 T	1,2,4-Trimethylbenzene	2.541	2.538	0.1	96	0.00
85 T	sec-Butylbenzene	3.143	3.053	2.9	93	0.00
86 T	p-Isopropyltoluene	2.745	2.725	0.7	95	0.00
87 T	1,3-Dichlorobenzene	1.592	1.586	0.4	98	0.00
88 T	1,4-Dichlorobenzene	1.629	1.584	2.8	98	0.00
89 T	n-Butylbenzene	2.113	2.143	-1.4	96	0.00
90 T	Hexachloroethane	0.467	0.453	3.0	94	0.00
91 T	1,2-Dichlorobenzene	1.629	1.630	-0.1	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.224	0.237	-5.8	99	0.00
93 T	1,2,4-Trichlorobenzene	1.067	1.124	-5.3	101	0.00
94 T	Hexachlorobutadiene	0.539	0.517	4.1	97	0.00
95 T	Naphthalene	3.289	3.537	-7.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061924\
Data File : VX041936.D
Acq On : 19 Jun 2024 10:21
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 00:54:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.125	1.181	-5.0	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6