

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046761.D
 Acq On : 19 Jun 2025 11:03
 Operator : JC/MD
 Sample : Q2345-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20250616

Quant Time: Jun 20 05:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209152	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82438	47.830	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.660%
35) Dibromofluoromethane	5.397	113	69253	50.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.080%
50) Toluene-d8	8.647	98	247834	49.849	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.700%
62) 4-Bromofluorobenzene	11.079	95	92518	49.921	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.840%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5018	8.731	ug/l	100
20) Methylene Chloride	2.800	84	2259	1.413	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046761.D
Acq On : 19 Jun 2025 11:03
Operator : JC/MD
Sample : Q2345-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20250616

Quant Time: Jun 20 05:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

