

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046767.D
 Acq On : 19 Jun 2025 13:15
 Operator : JC/MD
 Sample : Q2334-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GBTW1

Quant Time: Jun 20 05:19:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	164981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	283367	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	125261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110074	48.236	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.480%
35) Dibromofluoromethane	5.397	113	92057	49.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	8.653	98	334848	49.711	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	125274	49.892	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.780%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1511	1.986	ug/l #	77
17) Carbon Disulfide	2.532	76	4809	0.874	ug/l	98
64) Tetrachloroethene	9.275	164	2566	1.460	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046767.D
Acq On : 19 Jun 2025 13:15
Operator : JC/MD
Sample : Q2334-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GBTW1

Quant Time: Jun 20 05:19:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

