

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046769.D
 Acq On : 19 Jun 2025 13:57
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 20 05:20:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	170391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	287341	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	126696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109480	46.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.900%
35) Dibromofluoromethane	5.397	113	90683	47.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	8.653	98	343599	50.305	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	129390	50.819	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.640%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.489	56	4427	1.254	ug/l	90
39) Methylcyclohexane	7.385	83	8169	2.262	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046769.D
Acq On : 19 Jun 2025 13:57
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 20 05:20:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

