

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062019\
 Data File : VX010359.D
 Acq On : 20 Jun 2019 16:01
 Operator : JC/SP
 Sample : K3469-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-MW-3-8.0-061919

Quant Time: Jun 21 01:56:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	481083	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	417909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	145007	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	5.50	113	146502	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	8.71	98	550310	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	11.13	95	184107	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
Target Compounds						
16) Acetone	2.45	43	6510	4.423	ug/l	95
19) Methyl tert-butyl Ether	3.19	73	3783	0.443	ug/l #	90
78) n-propylbenzene	11.36	91	5757	0.401	ug/l	99
95) Naphthalene	13.83	128	16836	1.248	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062019\
Data File : VX010359.D
Acq On : 20 Jun 2019 16:01
Operator : JC/SP
Sample : K3469-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-MW-3-8.0-061919

Quant Time: Jun 21 01:56:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

