

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062019\
 Data File : VX010362.D
 Acq On : 20 Jun 2019 17:11
 Operator : JC/SP
 Sample : K3469-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-MW-7-10.0-061919

Quant Time: Jun 21 02:05:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	316201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	479022	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	424356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199725	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	145864	49.20	ug/l	0.00
Spiked Amount			Recovery	=	98.40%	
35) Dibromofluoromethane	5.50	113	145019	49.48	ug/l	0.00
Spiked Amount			Recovery	=	98.96%	
50) Toluene-d8	8.71	98	558019	49.77	ug/l	0.00
Spiked Amount			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	11.13	95	187830	46.40	ug/l	0.00
Spiked Amount			Recovery	=	92.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.04	59	4532	5.842	ug/l #	85
16) Acetone	2.45	43	12368	8.324	ug/l	97
19) Methyl tert-butyl Ether	3.20	73	17475	2.027	ug/l #	77
31) Cyclohexane	5.59	56	13205	3.084	ug/l	93
39) Methylcyclohexane	7.46	83	8966	1.792	ug/l	92
52) Toluene	8.79	92	2759	0.353	ug/l	88
68) m/p-Xylenes	10.36	106	5440	0.965	ug/l	99
73) Isopropylbenzene	11.02	105	89635	6.621	ug/l	100
78) n-propylbenzene	11.36	91	35904	2.413	ug/l	97
83) tert-Butylbenzene	11.77	119	24609	2.197	ug/l	90
85) sec-Butylbenzene	11.94	105	81037	6.161	ug/l	96
89) n-Butylbenzene	12.39	91	16757	1.621	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062019\
Data File : VX010362.D
Acq On : 20 Jun 2019 17:11
Operator : JC/SP
Sample : K3469-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-MW-7-10.0-061919

Quant Time: Jun 21 02:05:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

