

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029611.D
 Acq On : 20 Jun 2022 10:38
 Operator : JC/MD
 Sample : VX0620MBL01
 Misc : 5.00g10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620MBL01

Quant Time: Jun 21 01:47:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	277344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	464037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	393927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160868	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154330	42.037	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.080%
35) Dibromofluoromethane	5.385	113	136861	45.257	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	8.647	98	520353	47.101	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.200%
62) 4-Bromofluorobenzene	11.079	95	180947	43.433	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029611.D
Acq On : 20 Jun 2022 10:38
Operator : JC/MD
Sample : VX0620MBL01
Misc : 5.00g10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0620MBL01

Quant Time: Jun 21 01:47:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

