

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029615.D
 Acq On : 20 Jun 2022 12:10
 Operator : JC/MD
 Sample : N3325-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL

Quant Time: Jun 21 01:49:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456782	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	392554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166486	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154184	42.999	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.000%		
35) Dibromofluoromethane	5.391	113	134792	45.281	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.560%		
50) Toluene-d8	8.653	98	515596	47.412	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.820%		
62) 4-Bromofluorobenzene	11.079	95	182412	44.480	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.960%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	29113	29.080	ug/l #	92
16) Acetone	2.386	43	4575	3.201	ug/l	99
17) Carbon Disulfide	2.514	76	2710	0.381	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	33455	3.511	ug/l #	42
22) Diisopropyl ether	3.764	45	2330	0.266	ug/l #	81
31) Cyclohexane	5.477	56	17067	3.545	ug/l	91
39) Methylcyclohexane	7.385	83	10175	2.026	ug/l	99
40) Benzene	6.044	78	327551	27.427	ug/l	96
52) Toluene	8.720	92	8285	1.083	ug/l	97
67) Ethyl Benzene	10.195	91	51431	3.712	ug/l	99
68) m/p-Xylenes	10.305	106	7313	1.353	ug/l	100
69) o-Xylene	10.647	106	3970	0.744	ug/l	97
73) Isopropylbenzene	10.964	105	95490	7.650	ug/l	100
78) n-propylbenzene	11.305	91	99003	7.042	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	36926	3.522	ug/l	100
85) sec-Butylbenzene	11.890	105	6225	0.490	ug/l	97
89) n-Butylbenzene	12.335	91	7186	0.800	ug/l #	69
95) Naphthalene	13.774	128	7815	0.669	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029615.D
Acq On : 20 Jun 2022 12:10
Operator : JC/MD
Sample : N3325-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL

Quant Time: Jun 21 01:49:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

