

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029617.D
 Acq On : 20 Jun 2022 12:57
 Operator : JC/MD
 Sample : N3325-05DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16DL

Quant Time: Jun 21 01:49:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	269139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	396523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154061	43.243	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.480%		
35) Dibromofluoromethane	5.385	113	136324	45.865	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.740%		
50) Toluene-d8	8.653	98	516431	47.561	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.120%		
62) 4-Bromofluorobenzene	11.085	95	189339	46.239	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.480%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.958	59	13039	7.831	ug/l #	90
16) Acetone	2.385	43	2901	2.043	ug/l #	85
31) Cyclohexane	5.482	56	11926	2.493	ug/l	89
39) Methylcyclohexane	7.384	83	6245	1.245	ug/l	95
40) Benzene	6.043	78	232673	19.512	ug/l	98
52) Toluene	8.726	92	3450	0.452	ug/l	91
67) Ethyl Benzene	10.195	91	38705	2.765	ug/l	98
68) m/p-Xylenes	10.305	106	4493	0.823	ug/l	94
73) Isopropylbenzene	10.963	105	49891	3.775	ug/l	99
78) n-propylbenzene	11.304	91	128720	8.647	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	3517	0.317	ug/l	98
85) sec-Butylbenzene	11.890	105	11704	0.870	ug/l	94
89) n-Butylbenzene	12.335	91	10722	1.128	ug/l #	44
95) Naphthalene	13.780	128	6819	0.551	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029617.D
Acq On : 20 Jun 2022 12:57
Operator : JC/MD
Sample : N3325-05DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16DL

Quant Time: Jun 21 01:49:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

