

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029621.D
 Acq On : 20 Jun 2022 14:30
 Operator : JC/MD
 Sample : N3290-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19DL

Quant Time: Jun 21 01:50:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	268723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454585	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	388622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153185	43.064	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.120%
35) Dibromofluoromethane	5.391	113	133911	45.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.653	98	511218	47.237	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	182926	44.821	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.640%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	31587	32.705	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	239785	25.370	ug/l	97
22) Diisopropyl ether	3.758	45	2715	0.312	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029621.D
Acq On : 20 Jun 2022 14:30
Operator : JC/MD
Sample : N3290-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19DL

Quant Time: Jun 21 01:50:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

