

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029628.D
 Acq On : 20 Jun 2022 17:13
 Operator : JC/MD
 Sample : N3166-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB07-20220601

Quant Time: Jun 21 01:52:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	261004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	443934	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	380877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149304	43.214	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.420%		
35) Dibromofluoromethane	5.391	113	130630	45.153	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.300%		
50) Toluene-d8	8.653	98	504549	47.739	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.480%		
62) 4-Bromofluorobenzene	11.085	95	180675	45.331	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.660%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	18964	13.771	ug/l	97
20) Methylene Chloride	2.788	84	2322	0.757	ug/l	86
27) cis-1,2-Dichloroethene	4.501	96	1473	0.436	ug/l	91
44) Trichloroethene	7.129	130	4674	0.298	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029628.D
Acq On : 20 Jun 2022 17:13
Operator : JC/MD
Sample : N3166-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB07-20220601

Quant Time: Jun 21 01:52:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

