

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029638.D
 Acq On : 21 Jun 2022 04:12
 Operator : JC/MD
 Sample : N3354-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-46

Quant Time: Jun 21 08:47:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	266227	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	450580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149390	42.391	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.780%
35) Dibromofluoromethane	5.391	113	133651	45.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.040%
50) Toluene-d8	8.653	98	512353	47.762	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.085	95	185485	45.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.700%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3689	2.626	ug/l	95
17) Carbon Disulfide	2.514	76	2071	0.296	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029638.D
Acq On : 21 Jun 2022 04:12
Operator : JC/MD
Sample : N3354-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-46

Quant Time: Jun 21 08:47:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

