

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029641.D
 Acq On : 21 Jun 2022 05:22
 Operator : JC/MD
 Sample : N3354-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-49

Quant Time: Jun 21 08:47:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453548	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	387523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151676	42.746	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.500%
35) Dibromofluoromethane	5.391	113	132854	44.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.900%
50) Toluene-d8	8.653	98	507140	46.967	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.940%
62) 4-Bromofluorobenzene	11.079	95	176792	43.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.840%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11203	7.921	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	29424	3.121	ug/l	99
22) Diisopropyl ether	3.764	45	2474	0.285	ug/l #	77
73) Isopropylbenzene	10.970	105	2898	0.236	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029641.D
Acq On : 21 Jun 2022 05:22
Operator : JC/MD
Sample : N3354-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-49

Quant Time: Jun 21 08:47:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

