

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029643.D
 Acq On : 21 Jun 2022 06:09
 Operator : JC/MD
 Sample : N3354-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-50D

Quant Time: Jun 21 08:47:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	264996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	387967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152890	43.585	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.180%		
35) Dibromofluoromethane	5.385	113	134423	45.388	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.780%		
50) Toluene-d8	8.653	98	509482	47.089	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.180%		
62) 4-Bromofluorobenzene	11.079	95	180721	44.292	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.580%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	3301	2.361	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	118359	12.699	ug/l	98
22) Diisopropyl ether	3.764	45	4931	0.575	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029643.D
Acq On : 21 Jun 2022 06:09
Operator : JC/MD
Sample : N3354-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-50D

Quant Time: Jun 21 08:47:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

