

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029650.D
 Acq On : 21 Jun 2022 08:51
 Operator : JC/MD
 Sample : N3354-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-57

Quant Time: Jun 21 09:51:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454533	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153318	42.863	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.720%
35) Dibromofluoromethane	5.391	113	135072	45.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.200%
50) Toluene-d8	8.653	98	514992	47.591	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.180%
62) 4-Bromofluorobenzene	11.079	95	178951	43.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.700%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	2955	2.073	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029650.D
Acq On : 21 Jun 2022 08:51
Operator : JC/MD
Sample : N3354-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-57

Quant Time: Jun 21 09:51:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

