

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
 Data File : VX029653.D
 Acq On : 21 Jun 2022 10:01
 Operator : JC/MD
 Sample : N3354-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-60

Quant Time: Jun 21 17:09:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	250800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	425501	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	368912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146683	44.183	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.360%
35) Dibromofluoromethane	5.391	113	126341	45.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.120%
50) Toluene-d8	8.653	98	476892	47.077	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.079	95	172522	45.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	27498	29.860	ug/l	99
16) Acetone	2.380	43	10353	7.824	ug/l	89
19) Methyl tert-butyl Ether	3.117	73	77609	8.798	ug/l	98
40) Benzene	6.044	78	9195	0.827	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062022\
Data File : VX029653.D
Acq On : 21 Jun 2022 10:01
Operator : JC/MD
Sample : N3354-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-60

Quant Time: Jun 21 17:09:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

