

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036230.D
 Acq On : 20 Jun 2023 10:00
 Operator : JC/MD
 Sample : VX0620WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBL01

Quant Time: Jun 21 01:56:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	284976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	120747	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	155763	57.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.620%
35) Dibromofluoromethane	5.379	113	101693	52.623	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.240%
50) Toluene-d8	8.647	98	345436	48.868	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	139443	48.869	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036230.D
Acq On : 20 Jun 2023 10:00
Operator : JC/MD
Sample : VX0620WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0620WBL01

Quant Time: Jun 21 01:56:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

