

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036233.D
 Acq On : 20 Jun 2023 11:23
 Operator : JC/MD
 Sample : 03238-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Jun 21 01:58:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	114681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	153757	52.962	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.920%	
35) Dibromofluoromethane	5.385	113	100299	51.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.860%	
50) Toluene-d8	8.647	98	355141	49.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.560%	
62) 4-Bromofluorobenzene	11.079	95	135568	47.078	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.160%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2068	1.715	ug/l	90
30) Chloroform	5.099	83	5723	1.424	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036233.D
Acq On : 20 Jun 2023 11:23
Operator : JC/MD
Sample : 03238-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Jun 21 01:58:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

