

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036234.D
 Acq On : 20 Jun 2023 11:46
 Operator : JC/MD
 Sample : 03311-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1646

Quant Time: Jun 21 01:59:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	267128	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	110836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150416	59.579	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 119.160%#		
35) Dibromofluoromethane	5.385	113	96642	53.351	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.700%		
50) Toluene-d8	8.647	98	337050	50.867	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.740%		
62) 4-Bromofluorobenzene	11.079	95	129972	48.593	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 97.180%		
Target Compounds						
16) Acetone	2.386	43	3404	3.245	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036234.D
Acq On : 20 Jun 2023 11:46
Operator : JC/MD
Sample : 03311-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1646

Quant Time: Jun 21 01:59:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

