

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036235.D
 Acq On : 20 Jun 2023 12:09
 Operator : JC/MD
 Sample : 03274-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-061423

Quant Time: Jun 21 01:59:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	281086	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	151227	54.201	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.400%		
35) Dibromofluoromethane	5.379	113	95706	50.211	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.420%		
50) Toluene-d8	8.647	98	341521	48.982	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.960%		
62) 4-Bromofluorobenzene	11.079	95	130284	46.291	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.580%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036235.D
Acq On : 20 Jun 2023 12:09
Operator : JC/MD
Sample : 03274-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-061423

Quant Time: Jun 21 01:59:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

