

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036245.D
 Acq On : 20 Jun 2023 16:04
 Operator : JC/MD
 Sample : 03339-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10

Quant Time: Jun 21 08:11:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	160716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	294064	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	124057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151378	51.542	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.080%
35) Dibromofluoromethane	5.385	113	97841	49.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.647	98	362196	49.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	142507	48.399	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3888	3.186	ug/l	96
29) Tetrahydrofuran	5.013	42	8762	8.076	ug/l	98
31) Cyclohexane	5.470	56	6151	1.952	ug/l #	81
39) Methylcyclohexane	7.373	83	3203	1.012	ug/l #	26
40) Benzene	6.050	78	2310	0.280	ug/l	93
52) Toluene	8.720	92	9345	1.807	ug/l	99
67) Ethyl Benzene	10.195	91	27271	2.820	ug/l	96
68) m/p-Xylenes	10.299	106	37405	10.251	ug/l	96
69) o-Xylene	10.640	106	11285	3.135	ug/l	96
73) Isopropylbenzene	10.963	105	10927	1.187	ug/l	96
78) n-propylbenzene	11.305	91	8801	0.813	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	7575	0.966	ug/l	97
83) tert-Butylbenzene	11.713	119	2692	0.361	ug/l	85
84) 1,2,4-Trimethylbenzene	11.750	105	35483	4.538	ug/l	98
85) sec-Butylbenzene	11.890	105	11555	1.245	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036245.D
Acq On : 20 Jun 2023 16:04
Operator : JC/MD
Sample : 03339-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10

Quant Time: Jun 21 08:11:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

