

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036246.D
 Acq On : 20 Jun 2023 16:27
 Operator : JC/MD
 Sample : 03339-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jun 21 02:04:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	279680	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140398	51.053	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.100%
35) Dibromofluoromethane	5.385	113	93411	49.253	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	8.647	98	325309	46.892	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.780%
62) 4-Bromofluorobenzene	11.079	95	119556	42.693	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.380%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2356	2.062	ug/l	94
52) Toluene	8.720	92	6254	1.271	ug/l	99
67) Ethyl Benzene	10.195	91	9743	1.091	ug/l	91
68) m/p-Xylenes	10.299	106	13874	4.118	ug/l	97
69) o-Xylene	10.646	106	4021	1.210	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	11077	1.660	ug/l	95
95) Naphthalene	13.774	128	7531	1.118	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036246.D
Acq On : 20 Jun 2023 16:27
Operator : JC/MD
Sample : 03339-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jun 21 02:04:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

