

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036252.D
 Acq On : 20 Jun 2023 18:48
 Operator : JC/MD
 Sample : 03291-03DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6ADL

Quant Time: Jun 21 02:54:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262893	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	128768	48.956	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.920%
35) Dibromofluoromethane	5.385	113	87722	49.207	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	8.647	98	312341	47.897	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.079	95	116562	44.282	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.560%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	8964	3.177	ug/l	87
39) Methylcyclohexane	7.379	83	2842	1.005	ug/l	98
40) Benzene	6.037	78	399859	54.147	ug/l	98
52) Toluene	8.720	92	1642728	355.299	ug/l	99
67) Ethyl Benzene	10.195	91	219384	25.547	ug/l	99
68) m/p-Xylenes	10.299	106	244345	75.418	ug/l	100
69) o-Xylene	10.640	106	88933	27.822	ug/l	96
73) Isopropylbenzene	10.963	105	4830	0.618	ug/l	93
78) n-propylbenzene	11.305	91	15494	1.686	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	22185	3.336	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	92648	13.967	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036252.D
Acq On : 20 Jun 2023 18:48
Operator : JC/MD
Sample : 03291-03DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6ADL

Quant Time: Jun 21 02:54:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

