

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
 Data File : VX036253.D
 Acq On : 20 Jun 2023 19:12
 Operator : JC/MD
 Sample : 03339-02DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-08DL

Quant Time: Jun 21 02:54:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130989	58.140	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.280%	
35) Dibromofluoromethane	5.385	113	85878	52.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.000%	
50) Toluene-d8	8.647	98	293072	48.984	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.960%	
62) 4-Bromofluorobenzene	11.079	95	113600	47.037	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.080%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1185	0.699	ug/l #	93
39) Methylcyclohexane	7.379	83	1213	0.467	ug/l #	82
40) Benzene	6.044	78	12325	1.819	ug/l	97
52) Toluene	8.720	92	15489	3.651	ug/l	94
67) Ethyl Benzene	10.195	91	235752	29.617	ug/l	99
68) m/p-Xylenes	10.299	106	308430	102.700	ug/l	100
69) o-Xylene	10.640	106	147034	49.623	ug/l	99
73) Isopropylbenzene	10.963	105	11061	1.476	ug/l	98
78) n-propylbenzene	11.305	91	32561	3.694	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	56533	8.860	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	244343	38.391	ug/l	95
95) Naphthalene	13.774	128	80272	12.493	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062023\
Data File : VX036253.D
Acq On : 20 Jun 2023 19:12
Operator : JC/MD
Sample : 03339-02DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-08DL

Quant Time: Jun 21 02:54:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

