

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062024\
 Data File : VX041960.D
 Acq On : 20 Jun 2024 13:01
 Operator : JC/MD
 Sample : P2931-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240614022-01

Quant Time: Jun 21 03:16:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	30253	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	166332	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	144022	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	64579	31.009	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.367%
60) 4-Bromofluorobenzene	11.079	95	60098	26.958	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.867%
63) Toluene-d8	8.647	98	188600	30.608	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.033%
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	9434	29.476	ug/l	92
25) Chloroform	5.092	83	274620	81.963	ug/l	99
41) Bromodichloromethane	7.824	83	14740	6.456	ug/l	98
53) Dibromochloromethane	9.525	129	2355	1.161	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062024\
Data File : VX041960.D
Acq On : 20 Jun 2024 13:01
Operator : JC/MD
Sample : P2931-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240614022-01

Quant Time: Jun 21 03:16:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

