

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
 Data File : VX046792.D
 Acq On : 20 Jun 2025 18:42
 Operator : JC/MD
 Sample : VX0620WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02

Quant Time: Jun 21 00:12:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	120705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	197926	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94174	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	80412	48.163	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.320%
35) Dibromofluoromethane	5.397	113	65229	49.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	8.647	98	237880	50.560	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.079	95	92419	52.696	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.400%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.185	85	23492	18.229	ug/l	95
3) Chloromethane	1.307	50	24550	17.650	ug/l	98
4) Vinyl Chloride	1.386	62	27073	18.252	ug/l	93
5) Bromomethane	1.630	94	17332	20.769	ug/l	97
6) Chloroethane	1.703	64	16567	18.428	ug/l	99
7) Trichlorofluoromethane	1.904	101	40820	18.124	ug/l	100
8) Diethyl Ether	2.148	74	14298	18.498	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	25459	18.426	ug/l	99
10) Methyl Iodide	2.471	142	28127	19.187	ug/l	99
11) Tert butyl alcohol	2.953	59	9494	63.356	ug/l	97
12) 1,1-Dichloroethene	2.337	96	24747	18.561	ug/l	97
13) Acrolein	2.252	56	17936	82.173	ug/l	94
14) Allyl chloride	2.678	41	43381	17.958	ug/l	98
15) Acrylonitrile	3.075	53	56053	83.186	ug/l	99
16) Acetone	2.386	43	40885	73.434	ug/l	98
17) Carbon Disulfide	2.526	76	66786	16.588	ug/l	98
18) Methyl Acetate	2.715	43	23452	16.585	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	72719	17.517	ug/l	100
20) Methylene Chloride	2.806	84	28293	18.273	ug/l	96
21) trans-1,2-Dichloroethene	3.105	96	25885	18.145	ug/l	98
22) Diisopropyl ether	3.770	45	90685	19.531	ug/l	96
23) Vinyl Acetate	3.733	43	329720	89.301	ug/l	96
24) 1,1-Dichloroethane	3.623	63	48463	18.383	ug/l	99
25) 2-Butanone	4.568	43	61901	78.612	ug/l	99
26) 2,2-Dichloropropane	4.489	77	33595	17.473	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	30937	18.459	ug/l	99
28) Bromochloromethane	4.910	49	24187	20.238	ug/l	99
29) Tetrahydrofuran	5.013	42	39437	77.540	ug/l	99
30) Chloroform	5.105	83	50023	18.974	ug/l	100
31) Cyclohexane	5.489	56	46034	18.414	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	42194	18.249	ug/l	99
36) 1,1-Dichloropropene	5.702	75	34769	18.393	ug/l	98
37) Ethyl Acetate	4.727	43	28090	16.098	ug/l	99
38) Carbon Tetrachloride	5.684	117	37276	18.170	ug/l	98
39) Methylcyclohexane	7.385	83	45496	18.289	ug/l	97
40) Benzene	6.050	78	107324	19.184	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
 Data File : VX046792.D
 Acq On : 20 Jun 2025 18:42
 Operator : JC/MD
 Sample : VX0620WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02

Quant Time: Jun 21 00:12:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	17152	18.438	ug/l	93
42) 1,2-Dichloroethane	6.092	62	36923	18.781	ug/l	97
43) Isopropyl Acetate	6.348	43	47152	17.269	ug/l	99
44) Trichloroethene	7.135	130	27024	18.954	ug/l	94
45) 1,2-Dichloropropane	7.434	63	27100	19.546	ug/l	98
46) Dibromomethane	7.586	93	18189	18.254	ug/l	99
47) Bromodichloromethane	7.824	83	39394	19.373	ug/l	100
48) Methyl methacrylate	7.696	41	24127	17.619	ug/l	99
49) 1,4-Dioxane	7.665	88	5621	302.408	ug/l #	92
51) 4-Methyl-2-Pentanone	8.574	43	139590	85.717	ug/l	98
52) Toluene	8.720	92	67054	19.334	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	34782	18.435	ug/l	99
54) cis-1,3-Dichloropropene	8.372	75	39974	18.663	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	25004	19.346	ug/l	96
56) Ethyl methacrylate	9.116	69	34109	17.877	ug/l	99
57) 1,3-Dichloropropane	9.311	76	43248	19.431	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	96768	97.638	ug/l	99
59) 2-Hexanone	9.427	43	92283	81.968	ug/l	99
60) Dibromochloromethane	9.519	129	29076	19.089	ug/l	100
61) 1,2-Dibromoethane	9.610	107	25121	19.108	ug/l	100
64) Tetrachloroethene	9.275	164	22518	17.854	ug/l	98
65) Chlorobenzene	10.079	112	75459	18.778	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	25391	18.800	ug/l	99
67) Ethyl Benzene	10.195	91	131355	18.708	ug/l	98
68) m/p-Xylenes	10.299	106	99811	38.150	ug/l	99
69) o-Xylene	10.640	106	47475	19.363	ug/l	96
70) Styrene	10.653	104	81295	18.946	ug/l	99
71) Bromoform	10.799	173	17909	17.412	ug/l #	99
73) Isopropylbenzene	10.957	105	127763	18.421	ug/l	100
74) N-amyl acetate	10.842	43	42248	16.077	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	33761	17.433	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	33851	19.626	ug/l	96
77) Bromobenzene	11.195	156	29716	17.804	ug/l	98
78) n-propylbenzene	11.299	91	151361	18.373	ug/l	100
79) 2-Chlorotoluene	11.360	91	89555	18.207	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	104782	18.384	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	9153	15.000	ug/l	99
82) 4-Chlorotoluene	11.451	91	103304	17.996	ug/l	99
83) tert-Butylbenzene	11.713	119	105843	17.965	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	105636	18.447	ug/l	99
85) sec-Butylbenzene	11.890	105	137372	18.449	ug/l	100
86) p-Isopropyltoluene	12.006	119	113713	18.100	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	58189	17.880	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	58523	17.508	ug/l	99
89) n-Butylbenzene	12.329	91	108029	18.110	ug/l	100
90) Hexachloroethane	12.536	117	18545	16.857	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	55558	18.269	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	5776	15.107	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	37631	17.405	ug/l	98
94) Hexachlorobutadiene	13.719	225	15522	17.066	ug/l	99
95) Naphthalene	13.774	128	102000	16.559	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	36973	17.875	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
Data File : VX046792.D
Acq On : 20 Jun 2025 18:42
Operator : JC/MD
Sample : VX0620WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02

Quant Time: Jun 21 00:12:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
 Data File : VX046792.D
 Acq On : 20 Jun 2025 18:42
 Operator : JC/MD
 Sample : VX0620WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02

Quant Time: Jun 21 00:12:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:168 Resp: 120705
Ion Ratio Lower Upper
168 100
99 58.7 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 18.229 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 85 Resp: 23492
Ion Ratio Lower Upper
85 100
87 30.6 16.6 49.8









#3
Chloromethane
Concen: 17.650 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 50 Resp: 24550
Ion Ratio Lower Upper
50 100
52 32.5 24.9 37.3



#4
Vinyl Chloride
Concen: 18.252 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 62 Resp: 27073
Ion Ratio Lower Upper
62 100
64 33.6 23.9 35.9









#5
Bromomethane
Concen: 20.769 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 94 Resp: 17332
Ion Ratio Lower Upper
94 100
96 87.5 72.3 108.5



#6
Chloroethane
Concen: 18.428 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 64 Resp: 16567
Ion Ratio Lower Upper
64 100
66 34.4 27.1 40.7









#7
Trichlorofluoromethane
Concen: 18.124 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:101 Resp: 40820
Ion Ratio Lower Upper
101 100
103 64.6 51.8 77.8



#8
Diethyl Ether
Concen: 18.498 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 74 Resp: 14298
Ion Ratio Lower Upper
74 100
45 103.0 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.426 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02

Tgt Ion:101 Resp: 25459
Ion Ratio Lower Upper
101 100
85 43.3 35.4 53.0
151 75.9 61.2 91.8



#10
Methyl Iodide
Concen: 19.187 ug/l
RT: 2.471 min Scan# 227
Delta R.T. 0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:142 Resp: 28127
Ion Ratio Lower Upper
142 100
127 45.4 37.0 55.6
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 63.356 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 59 Resp: 9494
Ion Ratio Lower Upper
59 100
57 9.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 18.561 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 96 Resp: 24747
Ion Ratio Lower Upper
96 100
61 162.0 126.6 189.8
98 61.6 50.7 76.1





#13

Acrolein

Concen: 82.173 ug/l

RT: 2.252 min Scan# 191

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :





VX0620WBSD02

Tgt Ion: 56 Resp: 17936

Ion Ratio Lower Upper

56 100

55 73.5 55.2 82.8



#14

Allyl chloride

Concen: 17.958 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 41 Resp: 43381


Ion Ratio Lower Upper

41 100

39 64.8 53.6 80.4

76 33.3 26.8 40.2









#15
Acrylonitrile
Concen: 83.186 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 53 Resp: 56053
Ion Ratio Lower Upper
53 100
52 82.7 65.1 97.7
51 36.4 28.9 43.3



#16
Acetone
Concen: 73.434 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 43 Resp: 40885
Ion Ratio Lower Upper
43 100
58 29.5 24.5 36.7









#17
Carbon Disulfide
Concen: 16.588 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 76 Resp: 66786
Ion Ratio Lower Upper
76 100
78 8.8 7.5 11.3



#18
Methyl Acetate
Concen: 16.585 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 43 Resp: 23452
Ion Ratio Lower Upper
43 100
74 23.0 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 17.517 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 73 Resp: 72719
Ion Ratio Lower Upper
73 100
57 22.9 18.2 27.2



#20
Methylene Chloride
Concen: 18.273 ug/l
RT: 2.806 min Scan# 282
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 84 Resp: 28293
Ion Ratio Lower Upper
84 100
49 120.5 101.3 151.9
51 37.5 31.0 46.6
86 61.1 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 18.145 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 96 Resp: 25885
Ion Ratio Lower Upper
96 100
61 142.0 112.5 168.7
98 62.2 51.9 77.9



#22
Diisopropyl ether
Concen: 19.531 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 45 Resp: 90685
Ion Ratio Lower Upper
45 100
43 119.0 91.9 137.9
87 25.4 22.0 33.0
59 11.5 9.1 13.7









#23
 Vinyl Acetate
 Concen: 89.301 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02


Tgt Ion: 43 Resp: 329720
 Ion Ratio Lower Upper
 43 100
 86 11.8 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 18.383 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42

Tgt Ion: 63 Resp: 48463
 Ion Ratio Lower Upper
 63 100
 98 7.6 4.0 11.9
 100 4.2 2.2 6.6









#25
2-Butanone
Concen: 78.612 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02

Tgt Ion: 43 Resp: 61901
Ion Ratio Lower Upper
43 100
72 26.0 20.2 30.4



#26
2,2-Dichloropropane
Concen: 17.473 ug/l
RT: 4.489 min Scan# 558
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 77 Resp: 33595
Ion Ratio Lower Upper
77 100
97 23.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 18.459 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 96 Resp: 30937

Ion Ratio Lower Upper

96 100

61 146.1 0.0 291.0

98 64.7 0.0 126.8



#28

Bromochloromethane

Concen: 20.238 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 49 Resp: 24187


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 75.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 77.540 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 42 Resp: 39437

Ion Ratio Lower Upper

42 100

72 44.4 34.9 52.3

71 40.8 32.2 48.2



#30

Chloroform

Concen: 18.974 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion: 83 Resp: 50023

Ion Ratio Lower Upper

83 100

85 66.7 53.7 80.5









#31
Cyclohexane
Concen: 18.414 ug/l
RT: 5.489 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 56 Resp: 46034
Ion Ratio Lower Upper
56 100
69 31.2 24.8 37.2
84 82.7 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 18.249 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 97 Resp: 42194
Ion Ratio Lower Upper
97 100
99 63.8 51.3 76.9
61 46.6 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 48.163 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :





VX0620WBSD02

Tgt Ion: 65 Resp: 80412

Ion Ratio Lower Upper

65 100

67 53.1 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 114 Resp: 197926


Ion Ratio Lower Upper

114 100

63 23.6 0.0 44.2

88 16.9 0.0 33.2





#35

Dibromofluoromethane

Concen: 49.811 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion:113 Resp: 65229

Ion Ratio Lower Upper

113 100

111 103.1 82.0 123.0

192 19.1 15.3 22.9



#36

1,1-Dichloropropene

Concen: 18.393 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 75 Resp: 34769


Ion Ratio Lower Upper

75 100

110 35.0 17.2 51.5

77 31.3 24.2 36.4





#37

Ethyl Acetate

Concen: 16.098 ug/l

RT: 4.727 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 43 Resp: 28090

Ion Ratio Lower Upper

43 100

61 12.3 9.7 14.5

70 9.8 8.1 12.1



#38

Carbon Tetrachloride

Concen: 18.170 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 117 Resp: 37276


Ion Ratio Lower Upper

117 100

119 91.8 75.2 112.8

121 30.2 24.1 36.1









#39
Methylcyclohexane
Concen: 18.289 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:	83	Resp:	45496
Ion	Ratio	Lower	Upper
83	100		
55	82.5	63.0	94.4
98	49.9	38.8	58.2



#40
Benzene
Concen: 19.184 ug/l
RT: 6.050 min Scan# 814
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:	78	Resp:	107324
Ion	Ratio	Lower	Upper
78	100		
77	23.1	18.8	28.2









#41
Methacrylonitrile
Concen: 18.438 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02

Tgt Ion: 41 Resp: 17152




Ion	Ratio	Lower	Upper
41	100		
39	51.2	45.9	68.9
67	71.0	61.2	91.8
52	28.7	25.5	38.3




#42
1,2-Dichloroethane
Concen: 18.781 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 62 Resp: 36923

Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	17.8





#43

Isopropyl Acetate

Concen: 17.269 ug/l

RT: 6.348 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 43 Resp: 47152

Ion Ratio Lower Upper

43 100

61 18.9 15.3 22.9

87 11.6 10.0 15.0



#44

Trichloroethene

Concen: 18.954 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.006 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion:130 Resp: 27024

Ion Ratio Lower Upper

130 100

95 95.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.546 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :





VX0620WBSD02

Tgt Ion: 63 Resp: 27100

Ion Ratio Lower Upper

63 100

65 29.9 22.9 34.3



#46

Dibromomethane

Concen: 18.254 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion: 93 Resp: 18189


Ion Ratio Lower Upper

93 100

95 84.7 66.2 99.4

174 95.8 77.0 115.6









#47
Bromodichloromethane
Concen: 19.373 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 83 Resp: 39394
Ion Ratio Lower Upper
83 100
85 64.6 51.4 77.2
127 8.6 6.9 10.3



#48
Methyl methacrylate
Concen: 17.619 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 41 Resp: 24127
Ion Ratio Lower Upper
41 100
69 80.6 63.4 95.2
39 56.5 45.4 68.0





#49

1,4-Dioxane

Concen: 302.408 ug/l

RT: 7.665 min Scan# 1079

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 88 Resp: 5621

Ion Ratio Lower Upper

88 100

43 28.4 31.2 46.8#

58 70.3 54.7 82.1



#50

Toluene-d8

Concen: 50.560 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion: 98 Resp: 237880

Ion Ratio Lower Upper

98 100

100 67.5 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 85.717 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 43 Resp: 139590
Ion Ratio Lower Upper
43 100
58 37.2 30.8 46.2



#52
Toluene
Concen: 19.334 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 92 Resp: 67054
Ion Ratio Lower Upper
92 100
91 171.6 134.8 202.2









#53
 t-1,3-Dichloropropene
 Concen: 18.435 ug/l
 RT: 8.982 min Scan# 1195
 Delta R.T. 0.006 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02


Tgt Ion: 75 Resp: 34782
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.6 38.4



#54
 cis-1,3-Dichloropropene
 Concen: 18.663 ug/l
 RT: 8.372 min Scan# 1195
 Delta R.T. 0.006 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42

Tgt Ion: 75 Resp: 39974
 Ion Ratio Lower Upper
 75 100
 77 30.5 25.3 37.9
 39 48.2 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 19.346 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:	97	Resp:	25004
Ion	Ratio	Lower	Upper
97	100		
83	88.3	67.2	100.8
85	55.7	44.1	66.1
99	59.3	50.3	75.5



#56
Ethyl methacrylate
Concen: 17.877 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:	69	Resp:	34109
Ion	Ratio	Lower	Upper
69	100		
41	69.7	55.4	83.2
39	38.1	29.7	44.5





#57

1,3-Dichloropropane

Concen: 19.431 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :





VX0620WBSD02

Tgt Ion: 76 Resp: 43248

Ion Ratio Lower Upper

76 100

78 32.7 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 97.638 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion: 63 Resp: 96768

Ion Ratio Lower Upper

63 100

106 28.0 22.6 34.0









#59
2-Hexanone
Concen: 81.968 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 43 Resp: 92283
Ion Ratio Lower Upper
43 100
58 54.1 26.8 80.4



#60
Dibromochloromethane
Concen: 19.089 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 129 Resp: 29076
Ion Ratio Lower Upper
129 100
127 77.1 38.4 115.2









#61
1,2-Dibromoethane
Concen: 19.108 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:107 Resp: 25121
Ion Ratio Lower Upper
107 100
109 94.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 52.696 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 95 Resp: 92419
Ion Ratio Lower Upper
95 100
174 75.1 0.0 150.4
176 71.5 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:117 Resp: 182045
Ion Ratio Lower Upper
117 100
82 55.3 44.6 66.8
119 32.6 25.8 38.8



#64
Tetrachloroethene
Concen: 17.854 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:164 Resp: 22518
Ion Ratio Lower Upper
164 100
166 132.2 104.6 157.0
129 97.7 76.3 114.5
131 95.5 74.1 111.1









#65
Chlorobenzene
Concen: 18.778 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:112 Resp: 75459
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 18.800 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:131 Resp: 25391
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.8
119 63.1 32.1 96.3





#67

Ethyl Benzene

Concen: 18.708 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.006 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :





VX0620WBSD02

Tgt Ion: 91 Resp: 131355

Ion Ratio Lower Upper

91 100

106 29.3 24.2 36.2



#68

m/p-Xylenes

Concen: 38.150 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion: 106 Resp: 99811

Ion Ratio Lower Upper

106 100

91 204.1 164.8 247.2









#69
o-Xylene
Concen: 19.363 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:106 Resp: 47475
Ion Ratio Lower Upper
106 100
91 215.6 111.3 333.9



#70
Styrene
Concen: 18.946 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:104 Resp: 81295
Ion Ratio Lower Upper
104 100
78 51.2 41.4 62.2
103 54.6 43.1 64.7









#71
Bromoform
Concen: 17.412 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:173 Resp: 17909
Ion Ratio Lower Upper
173 100
175 47.6 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:152 Resp: 94174
Ion Ratio Lower Upper
152 100
115 69.9 43.2 129.6
150 163.0 0.0 346.8









#73
Isopropylbenzene
Concen: 18.421 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:105 Resp: 127763
Ion Ratio Lower Upper
105 100
120 25.9 13.0 39.0



#74
N-ethyl acetate
Concen: 16.077 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 43 Resp: 42248
Ion Ratio Lower Upper
43 100
70 41.6 33.4 50.0
55 24.7 19.1 28.7
61 24.1 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 17.433 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:	83	Resp:	33761
Ion	Ratio	Lower	Upper
83	100		
131	10.6	4.9	14.6
85	64.9	32.0	96.0



#76
1,2,3-Trichloropropane
Concen: 19.626 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:	75	Resp:	33851
Ion	Ratio	Lower	Upper
75	100		
77	34.6	18.5	55.5









#77
Bromobenzene
Concen: 17.804 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:156 Resp: 29716
Ion Ratio Lower Upper
156 100
77 168.2 82.8 248.6
158 98.0 47.6 142.9



#78
n-propylbenzene
Concen: 18.373 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 91 Resp: 151361
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1









#79
2-Chlorotoluene
Concen: 18.207 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 91 Resp: 89555
Ion Ratio Lower Upper
91 100
126 32.8 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 18.384 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 105 Resp: 104782
Ion Ratio Lower Upper
105 100
120 48.9 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 15.000 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion: 75 Resp: 9153
Ion Ratio Lower Upper
75 100
53 108.9 86.8 130.2
89 43.6 36.9 55.3



#82
4-Chlorotoluene
Concen: 17.996 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 91 Resp: 103304
Ion Ratio Lower Upper
91 100
126 29.5 14.4 43.2









#83
tert-Butylbenzene
Concen: 17.965 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:119 Resp: 105843
Ion Ratio Lower Upper
119 100
91 58.6 29.5 88.6
134 23.5 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 18.447 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:105 Resp: 105636
Ion Ratio Lower Upper
105 100
120 43.7 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 18.449 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBSD02

Tgt Ion:105 Resp: 137372
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 18.100 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046792.D
 Acq: 20 Jun 2025 18:42

Tgt Ion:119 Resp: 113713
 Ion Ratio Lower Upper
 119 100
 134 26.4 12.9 38.7
 91 24.7 12.3 36.9









#87
1,3-Dichlorobenzene
Concen: 17.880 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:146 Resp: 58189
Ion Ratio Lower Upper
146 100
111 40.4 20.6 61.8
148 63.2 31.6 94.8



#88
1,4-Dichlorobenzene
Concen: 17.508 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:146 Resp: 58523
Ion Ratio Lower Upper
146 100
111 42.1 20.3 61.0
148 64.2 32.1 96.5





#89

n-Butylbenzene

Concen: 18.110 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion: 91 Resp: 108029

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3

134 25.4 12.9 38.6



#90

Hexachloroethane

Concen: 16.857 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046792.D




Acq: 20 Jun 2025 18:42


Tgt Ion: 117 Resp: 18545

Ion Ratio Lower Upper

117 100

201 69.9 34.4 103.3









#91
1,2-Dichlorobenzene
Concen: 18.269 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42




Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02


Tgt Ion:146 Resp: 55558
Ion Ratio Lower Upper
146 100
111 43.2 21.7 65.1
148 65.7 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 15.107 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion: 75 Resp: 5776
Ion Ratio Lower Upper
75 100
155 85.0 41.9 125.7
157 107.0 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 17.405 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBSD02





Tgt Ion:180 Resp: 37631

Ion Ratio Lower Upper

180 100

182 95.3 46.7 140.1

145 31.8 15.8 47.4



#94

Hexachlorobutadiene

Concen: 17.066 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046792.D

Acq: 20 Jun 2025 18:42




Tgt Ion:225 Resp: 15522


Ion Ratio Lower Upper

225 100

223 63.9 32.5 97.5

227 64.6 32.2 96.6









#95
Naphthalene
Concen: 16.559 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Instrument :
MSVOA_X
ClientSampleId :
VX0620WBSD02

Tgt Ion:128 Resp: 102000
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 17.875 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046792.D
Acq: 20 Jun 2025 18:42

Tgt Ion:180 Resp: 36973
Ion Ratio Lower Upper
180 100
182 94.6 47.8 143.4
145 33.3 17.3 51.9

