

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
 Data File : VX046787.D
 Acq On : 20 Jun 2025 16:58
 Operator : JC/MD
 Sample : Q2258-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 21 00:10:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	85297	48.216	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.440%
35) Dibromofluoromethane	5.397	113	70938	49.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.460%
50) Toluene-d8	8.653	98	260441	50.308	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.079	95	96949	50.238	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062025\
Data File : VX046787.D
Acq On : 20 Jun 2025 16:58
Operator : JC/MD
Sample : Q2258-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 21 00:10:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

