

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062119\
 Data File : VX010375.D
 Acq On : 21 Jun 2019 11:19
 Operator : JC/SP
 Sample : VX0621WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0621WBL01

Quant Time: Jun 21 23:24:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	287797	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	435816	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	371604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165653	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	142066	52.65	ug/l	0.00
Spiked Amount			Recovery	=	105.30%	
35) Dibromofluoromethane	5.49	113	142920	53.60	ug/l	0.00
Spiked Amount			Recovery	=	107.20%	
50) Toluene-d8	8.71	98	540433	52.98	ug/l	0.00
Spiked Amount			Recovery	=	105.96%	
62) 4-Bromofluorobenzene	11.13	95	173702	47.16	ug/l	0.00
Spiked Amount			Recovery	=	94.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062119\
Data File : VX010375.D
Acq On : 21 Jun 2019 11:19
Operator : JC/SP
Sample : VX0621WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0621WBL01

Quant Time: Jun 21 23:24:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

