

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062119\
 Data File : VX010377.D
 Acq On : 21 Jun 2019 12:18
 Operator : JC/SP
 Sample : K3476-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190620

Quant Time: Jun 21 23:31:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	305936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	461764	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	409828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192876	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	139244	48.54	ug/l	0.00
Spiked Amount						
			Recovery	=	97.08%	
35) Dibromofluoromethane	5.50	113	141812	50.19	ug/l	0.00
Spiked Amount						
			Recovery	=	100.38%	
50) Toluene-d8	8.71	98	542228	50.17	ug/l	0.00
Spiked Amount						
			Recovery	=	100.34%	
62) 4-Bromofluorobenzene	11.13	95	180605	46.28	ug/l	0.00
Spiked Amount						
			Recovery	=	92.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062119\
Data File : VX010377.D
Acq On : 21 Jun 2019 12:18
Operator : JC/SP
Sample : K3476-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190620

Quant Time: Jun 21 23:31:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

