

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062119\
 Data File : VX010397.D
 Acq On : 21 Jun 2019 20:05
 Operator : JC/SP
 Sample : K3486-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-3

Quant Time: Jun 24 16:32:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	289668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	451920	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	399899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	136913	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
35) Dibromofluoromethane	5.50	113	137254	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
50) Toluene-d8	8.71	98	528712	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.13	95	174747	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
Target Compounds						
11) Tert butyl alcohol	3.04	59	3418	4.810	ug/l #	85
16) Acetone	2.45	43	10001	7.348	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062119\
Data File : VX010397.D
Acq On : 21 Jun 2019 20:05
Operator : JC/SP
Sample : K3486-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-3

Quant Time: Jun 24 16:32:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

