

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062119\
 Data File : VX010392.D
 Acq On : 21 Jun 2019 18:08
 Operator : JC/SP
 Sample : K3477-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 22 00:54:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	293695	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	450274	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	404169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	189585	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137034	49.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.52%	
35) Dibromofluoromethane	5.50	113	137065	49.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.50%	
50) Toluene-d8	8.71	98	531696	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	178883	47.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.02%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062119\
Data File : VX010392.D
Acq On : 21 Jun 2019 18:08
Operator : JC/SP
Sample : K3477-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 22 00:54:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX010392.D

Acq: 21 Jun 2019 18:08

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SOIL-REF-6

Tot Ion:168 Resp: 293695

Ion Ratio Lower Upper

168 100

99 47.6 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

120000

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 49.764 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.01 min

Lab File: VX010392.D


Acq: 21 Jun 2019 18:08

Tgt Ion: 65 Resp: 137034

Ion Ratio Lower Upper

65 100

67 55.0 0.0 110.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.86 min Scan# 936
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tot Ion:114 Resp: 450274
Ion Ratio Lower Upper
114 100
63 17.0 0.0 33.8
88 14.6 0.0 27.6



#35
Dibromofluoromethane
Concen: 49.752 ug/l
RT: 5.50 min Scan# 713
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08

Tgt Ion:113 Resp: 137065
Ion Ratio Lower Upper
113 100
111 102.6 81.2 121.8
192 22.9 18.6 27.8









#50
Toluene-d8
Concen: 50.446 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tot Ion: 98 Resp: 531696
Ion Ratio Lower Upper
98 100
100 63.8 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 47.010 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08

Tgt Ion: 95 Resp: 178883
Ion Ratio Lower Upper
95 100
174 94.2 0.0 187.6
176 90.9 0.0 184.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tot Ion:117 Resp: 404169
Ion Ratio Lower Upper
117 100
82 52.0 41.2 61.8
119 32.3 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX010392.D
Acq: 21 Jun 2019 18:08

Tgt Ion:152 Resp: 189585
Ion Ratio Lower Upper
152 100
115 54.2 38.6 115.7
150 157.0 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

