

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022782.D
 Acq On : 21 Jun 2021 10:09
 Operator : JC/MD
 Sample : VX0621MBL01
 Misc : 5.0g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0621MBL01

Quant Time: Jun 22 01:09:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	114531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50159	48.402	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.800%
35) Dibromofluoromethane	5.397	113	36024	49.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	139727	48.506	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.020%
62) 4-Bromofluorobenzene	11.085	95	45637	41.890	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022782.D
Acq On : 21 Jun 2021 10:09
Operator : JC/MD
Sample : VX0621MBL01
Misc : 5.0g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0621MBL01

Quant Time: Jun 22 01:09:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

