

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022786.D
 Acq On : 21 Jun 2021 11:52
 Operator : JC/MD
 Sample : M2753-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20210616

Quant Time: Jun 22 01:11:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56559	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	113260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53362	52.460	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.920%	
35) Dibromofluoromethane	5.397	113	38271	52.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.920%	
50) Toluene-d8	8.653	98	151605	53.220	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.440%	
62) 4-Bromofluorobenzene	11.085	95	50265	46.656	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.320%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1770	3.886	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022786.D
Acq On : 21 Jun 2021 11:52
Operator : JC/MD
Sample : M2753-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20210616

Quant Time: Jun 22 01:11:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

