

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022787.D
 Acq On : 21 Jun 2021 12:16
 Operator : JC/MD
 Sample : M2753-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210616

Quant Time: Jun 22 01:11:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	118551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	106086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	39619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54587	51.643	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.280%	
35) Dibromofluoromethane	5.397	113	40468	53.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.000%	
50) Toluene-d8	8.653	98	158738	53.237	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.480%	
62) 4-Bromofluorobenzene	11.085	95	54475	48.307	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.620%	
Target Compounds						
16) Acetone	2.392	43	2013	4.253	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022787.D
Acq On : 21 Jun 2021 12:16
Operator : JC/MD
Sample : M2753-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210616

Quant Time: Jun 22 01:11:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

