

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022790.D
 Acq On : 21 Jun 2021 13:27
 Operator : JC/MD
 Sample : M2753-01DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210616DL

Quant Time: Jun 22 01:12:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55519	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	111182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52582	52.66	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.32%	
35) Dibromofluoromethane	5.397	113	37991	53.56	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.12%	
50) Toluene-d8	8.653	98	147893	52.89	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.78%	
62) 4-Bromofluorobenzene	11.085	95	49665	46.96	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.92%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.501	96	7951	8.74	ug/l	89
44) Trichloroethene	7.135	130	62291	80.51	ug/l	99
64) Tetrachloroethene	9.275	164	4505	7.37	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022790.D
Acq On : 21 Jun 2021 13:27
Operator : JC/MD
Sample : M2753-01DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20210616DL

Quant Time: Jun 22 01:12:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

