

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022791.D
 Acq On : 21 Jun 2021 13:50
 Operator : JC/MD
 Sample : M2793-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1

Quant Time: Jun 22 01:12:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53187	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	107784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	48376	50.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.140%
35) Dibromofluoromethane	5.397	113	33845	49.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	8.653	98	134241	49.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.040%
62) 4-Bromofluorobenzene	11.085	95	43846	42.765	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.540%
Target Compounds						
16) Acetone	2.386	43	930	2.171	ug/l	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022791.D
Acq On : 21 Jun 2021 13:50
Operator : JC/MD
Sample : M2793-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1

Quant Time: Jun 22 01:12:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

