

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022793.D
 Acq On : 21 Jun 2021 14:37
 Operator : JC/MD
 Sample : M2783-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20210617

Quant Time: Jun 22 01:13:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53915	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37386	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51914	53.539	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.080%
35) Dibromofluoromethane	5.397	113	37453	52.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.600%
50) Toluene-d8	8.653	98	146621	51.934	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.860%
62) 4-Bromofluorobenzene	11.085	95	50249	47.061	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022793.D
Acq On : 21 Jun 2021 14:37
Operator : JC/MD
Sample : M2783-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20210617

Quant Time: Jun 22 01:13:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

