

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022794.D
 Acq On : 21 Jun 2021 15:01
 Operator : JC/MD
 Sample : M2783-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20210617

Quant Time: Jun 22 01:13:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55235	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	107910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36731	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	52402	52.751	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.500%
35) Dibromofluoromethane	5.397	113	36171	52.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.080%
50) Toluene-d8	8.653	98	146087	53.826	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.660%
62) 4-Bromofluorobenzene	11.085	95	48933	47.671	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.340%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1237	2.781	ug/l	# 85
44) Trichloroethene	7.135	130	4721	6.287	ug/l	89
64) Tetrachloroethene	9.281	164	545	0.891	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022794.D
Acq On : 21 Jun 2021 15:01
Operator : JC/MD
Sample : M2783-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20210617

Quant Time: Jun 22 01:13:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

